

LIQUIZYME

BILIRUBIN

(DMSO Method)



Code	Product Name	Pack Size
LS009D	Liquizyme Bilirubin	240 ml
LS009E	Liquizyme Bilirubin	1200 ml

Intended Use

Diagnostic reagent for quantitative *in vitro* determination of Bilirubin in human serum.

Clinical Significance

Bilirubin is a breakdown product of haemoglobin. Bilirubin formed in the reticulo endothelial system is transported bound by albumin to the liver. This bilirubin is water insoluble and is known as indirect or unconjugated bilirubin. In the liver, bilirubin is conjugated to glucuronic acid to form direct bilirubin. Conjugated bilirubin is excreted via the biliary system into the intestine. Here it is metabolised by bacteria to urobilinogen & stercobilinogen.

Principle

In the determination of Bilirubin Total, Bilirubin is coupled with diazotized sulfanilic acid in the presence of ethylene glycol and dimethylsulfoxide as solvents to produce an intensely colored diazo dye. The intensity of colour of this solution is proportional to the concentration of the bilirubin total in the sample.

Reaction

Total Bilirubin

Bilirubin + Sulphanilic acid + Sodium Nitrite $\xrightarrow{\text{DMSO}}$ Azobilirubin

Direct Bilirubin

Bilirubin + Sulphanilic acid + Sodium Nitrite $\xrightarrow{\text{DMSO}}$ Azobilirubin

Contents:

Reagent 1 : Total Bilirubin Reagent

Buffer : < 15 mmol/l
Sulphanilic Acid : >20 mmol/l

Reagent 2 : Total Nitrite Reagent

Sodium Nitrite : > 10 mmol/l

Reagent 3 : Direct Bilirubin Reagent

Buffer : < 15 mmol/l
Sulphanilic Acid : >20 mmol/l

Reagent 4 : Direct Nitrite Reagent

Sodium Nitrite : < 10 mmol/l

Reagent 5: Bilirubin Artificial Standard : 10 mg/dl

Ready to use

Unit Conversion

mg/dl x 16.95 = μ mol/l

Normal Value :

Serum Total Bilirubin : upto 1.0 mg/dl
Direct Bilirubin : upto 0.3 mg/dl

Each Laboratory should establish its own normal range representing its patient population.

Total Bilirubin

Performance Data

Data contained within this section is representative of performance on Beacon system. Data obtained in your laboratory may differ from these values.

Limit of quantification : 0.0052 mg

Linearity : 20 mg/dl

Measuring range : 0.0052 – 20 mg/dl

Intra-assay precision Within run (n=20)	Mean (mg/dl)	SD (mg/dl)	CV (%)
Sample 1	1.06	0.03	3.16
Sample 2	4.47	0.04	0.92

Inter-assay precision Run to run (n=20)	Mean (mg/dl)	SD (mg/dl)	CV (%)
Sample 1	4.06	0.01	0.35

Comparison

A comparison between Beacon Bilirubin (Total & Direct) (y) and a commercially available test (x) using 20 samples gave following results:

y = 0.990 x + 0.042 mg/dl

r = 0.999

Direct Bilirubin

Performance Data

Data contained within this section is representative of performance on Beacon system. Data obtained in your laboratory may differ from these values.

Limit of quantification : 0.0052 mg

Linearity : 20 mg/dl

Measuring range : 0.0052 – 20 mg/dl

Intra-assay precision Within run (n=20)	Mean (mg/dl)	SD (mg/dl)	CV (%)
Sample 1	0.251	0.01	3.59
Sample 2	1.15	0.01	0.47

Inter-assay precision Run to run (n=20)	Mean (mg/dl)	SD (mg/dl)	CV (%)
Sample 1	1.16	0.01	1.02

Comparison

A comparison between Beacon Bilirubin (Total & Direct) (y) and a commercially available test (x) using 20 samples gave following results :

$$y = 0.993x + 0.011 \text{ mg/dl}$$

$$r = 0.999$$

Linearity:

This procedure is linear upto 20 mg/dl. If the values exceed this limit, dilute the sample with normal saline (NaCl 0.9%) and repeat the assay. Multiply result by dilution factor.

Interferences

Following substances do not interfere:
haemoglobin up to 7.5 g/l, triglycerides up to 1500 mg/dl.

Warning And Precautions

For in vitro diagnostic use. To be handled by entitled and professionally educated person.

Procedure:

Pipette into clean dry test tubes labeled as Blank (B), and Test (T) :

Total Bilirubin

Addition Sequence	(B)	(T)
Total Bilirubin Reagent	1000 µl	1000 µl
Total Bilirubin Activator	-	20 µl
Serum	50 µl	50 µl

Direct Bilirubin

Addition Sequence	(B)	(T)
Direct Bilirubin Reagent	1000 µl	1000 µl
Direct Bilirubin Activator	-	20 µl
Serum	50 µl	50 µl

Mix well and incubate at 37°C for exactly 5 minutes. Measure the absorbance of the Test Samples (Abs. T) immediately against their respective Blanks.

Calculation

With factor:

Total Bilirubin = OD of test - OD of sample blank x Factor (20)

With artificial standard:

$$\text{Bilirubin Concentration (mg/dl)} = \frac{\text{OD test} - \text{OD of sample Blank}}{\text{OD of standard}} \times 10$$

Assay Parameters For Photometers

Mode	End Point
Wavelength	546
Sample Volume (µl)	50
Total Bili / Direct Bili Reagent (µl)	1000
Total Bili / Direct Bili Activator (µl)	20
Incubation Time	5 min
Reaction temperature (°C)	37
Linearity Low (mg/dl)	0.0052
Linearity High (mg/dl)	20
Blank with	Serum
Standard Concentration	10 mg/dl
Unit	mg/dl

References

1. Cornall, A. G., Bardawill, C. J., David, M. M.: J. Biol. Chem. 177, 751, 1949.
2. Doumas, B. T., Bayse, D. D. a kol.: Clin. Chem. 27, 1642, 1981.
3. Chromý, V., Fischer, J.: Clin. Chem. 23, 754, 1977.
4. Chromý, V., Fischer, J., Vozniček, J.: Z. Med. Labor. Diagn. 21, 333, 1980.
5. Tietz Textbook of Clinical Chemistry and Molecular diagnostics. Burtis, C.A.
6. Ashwood, E.R., Bruns, D.E.; 5th edition, WB Saunders.

Symbols Used On Labels



Catalogue
Number



Manufacturer



See Instruction
for Use



Lot Number



Content



Storage Temperature



Expiry Date



In Vitro Diagnostics

